

Anorectal manometry 판독

1. The peak resting anal sphincter pressure
: 정상 40-100mmHg -> 증가 또는 감소시 anal dysfunction
 2. The squeezing pressure during maximal voluntary effort
정상 -> Ext.anal sphincter 가 resting 에 비해 2배 이상 압력증가
(절대수치는 70~240mmHg)
감소 -> resting 에 비해 2배미만 증가
 3. The rectoanal inhibitory response was present/absent
- 그래프보고 RAIR 박스를 확인, 박스내에 oscillation 있으면 RAIR+
 4. The rectal sensory threshold : normal 100-300cc, abnormal : 기준이하에서 요의를 느낄 때
 5. In straining, anal sphincter was paradoxically contracted (yes/no) : no 여야 정상
- 6개중 1개라도 이상 있으면 anal dysfunction 으로 적고 따로 기재(모두 정상이면 normal)
- Ballon expulsion test : 배출 실패시 -> defecatory dysfunction

3분 이내 배출시 -> normal

Duration of sustained squeeze : 12초가 정상, 이상이면 비정상

